

इंटरनेट

मानक

Disclosure to Promote the Right To Information

Whereas the Parliament of India has set out to provide a practical regime of right to information for citizens to secure access to information under the control of public authorities, in order to promote transparency and accountability in the working of every public authority, and whereas the attached publication of the Bureau of Indian Standards is of particular interest to the public, particularly disadvantaged communities and those engaged in the pursuit of education and knowledge, the attached public safety standard is made available to promote the timely dissemination of this information in an accurate manner to the public.

“जानने का अधिकार, जीने का अधिकार”

Mazdoor Kisan Shakti Sangathan


“The Right to Information, The Right to Live”

“पुराने को छोड़ नये के तरफ”

Jawaharlal Nehru

“Step Out From the Old to the New”

IS 6767-3 (1972): Direct Reading Pointer Indicator Type
ac/dc Electronic Voltmeter, Part 3: Up to 300 MHz [LITD 8:
Electronic Measuring Instruments, Systems and Accessories]



“ज्ञान से एक नये भारत का निर्माण”

Satyanarayan Gangaram Pitroda

“Invent a New India Using Knowledge”



“ज्ञान एक ऐसा खजाना है जो कभी चुराया नहीं जा सकता है”

Bhartrhari—Nitiśatakam

“Knowledge is such a treasure which cannot be stolen”

BLANK PAGE



IS : 6767 (Part III) - 1972

Indian Standard

SPECIFICATION FOR DIRECT READING POINTER INDICATOR TYPE AC/DC ELECTRONIC VOLTMETER

PART III UP TO 300 MHz

UDC 621.317.725 : 621.38



© Copyright 1973

INDIAN STANDARDS INSTITUTION
MANAK BHAVAN, 9 BAHADUR SHAH ZAFAR MARG
NEW DELHI 1

Price Rs 4.00

Gr. 3

March 1973

Indian Standard

SPECIFICATION FOR DIRECT READING POINTER INDICATOR TYPE AC/DC ELECTRONIC VOLTMETER

PART III UP TO 300 MHz

Electronic Equipment Sectional Committee, ETDC 24

Chairman

*SHRI Y. VENKATARAMIAH

Representing

Directorate General of All India Radio (Ministry of
Information & Broadcasting)

Members

- | | |
|--|---|
| <p>RESEARCH ENGINEER, ALL INDIA RADIO (<i>Alternate to</i>
Shri Y. Venkataramiah)</p> <p>ADDITIONAL DIRECTOR
DEPUTY DIRECTOR STANDARDS
(TELECOMMUNICATIONS) (<i>Alternate</i>)</p> <p>SHRI H. K. L. ARORA
SHRI R. G. KESWANI (<i>Alternate</i>)
(Bombay)</p> <p>SHRI ARUP CHAUDHURI (<i>Alternate</i>)
(Calcutta)</p> <p>SHRI L. S. V. EASWAR (<i>Alternate</i>)
(Madras)</p> <p>ASSISTANT DIRECTOR, ELECTRICAL Naval Headquarters
ENGINEERING (DESIGN)</p> <p>ASSISTANT DIRECTOR, ELECTRICAL
ENGINEERING (MATERIAL) (<i>Alternate</i>)</p> <p>SHRI H. V. BADRINATH</p> <p>SHRI A. M. JOSHI (<i>Alternate</i>)</p> <p>SHRI BALRAJ BHANOT</p> <p>BRIG B. BHASIN</p> <p>SHRI J. M. BORDE</p> <p>SHRI N. GANESAN (<i>Alternate</i>)</p> | <p>Railway Board (Ministry of Railways)</p> <p>All India Radio and Electronics Association, Bombay</p> <p>Wireless Planning & Co-ordination Wing (Department of Communications)</p> <p>Directorate General of Technical Development
(Ministry of Supply, Technical Development &
Materials Planning)</p> <p>Directorate General of Inspection (Ministry of
Defence)</p> <p>All India Instrument Manufacturers' and Dealers'
Association, Bombay</p> |
|--|---|

(Continued on page 2)

*Shri Y. Venkataramiah was Chairman for the meeting in which this standard was finalized.

© Copyright 1973

INDIAN STANDARDS INSTITUTION

This publication is protected under the *Indian Copyright Act* (XIV of 1957) and reproduction in whole or in part by any means except with written permission of the publisher shall be deemed to be an infringement of copyright under the said Act.

(Continued from page 1)

<i>Members</i>	<i>Representing</i>
DEPUTY DIRECTOR GENERAL	Overseas Communications Service (Department of Communications)
DIRECTOR (Alternate)	
SHRI B. P. GHOSH	National Test House, Calcutta
SHRI S. T. KAGALI	Bharat Electronics Ltd, Bangalore
SHRI K. S. KELKAR (Alternate I)	
SHRI R. S. MUNDKUR (Alternate II)	
WG CDR J. S. LAMBA	Directorate of Technical Development & Production (Air) (Ministry of Defence)
SHRI A. V. RAJU, SSO II (Alternate)	
COL N. S. MATHUR	Police Wireless (Ministry of Home Affairs)
SHRI R. S. KALE (Alternate)	
SHRI K. P. P. NAMBIAR	Radio Electronics & Television Manufacturers' Association (RETMA), Bombay
SHRI R. K. JAIN (Alternate)	
SHRI D. V. PHATAK	Radio Transmitters Subcommittee, ETDC 24 : 4, ISI
SHRI T. V. RAMAMURTI	Safety for Electronic Equipment Subcommittee, ETDC 24 : 3, ISI
SHRI V. RAMA SUBRAMANYAM	Directorate General of Civil Aviation (Department of Civil Aviation)
DR RAM PARSHAD	National Physical Laboratory (CSIR), New Delhi; and Electronic Measuring Equipment Subcommittee, ETDC 24 : 2, ISI
SHRI D. K. SACHDEV	Indian Telephone Industries Ltd, Bangalore
SHRI G. B. DIXIT (Alternate)	
SHRI M. SANKARALINGAM	Directorate General of Supplies & Disposals (Ministry of Supply, Technical Development & Materials Planning)
SHRI P. T. KRISHNAMCHAR (Alternate)	
SHRI T. V. SRIRANGAN	Posts & Telegraphs Board, New Delhi
SHRI N. SRINIVASAN,	Director General, ISI (Ex-officio Member)
Deputy Director (Elec tech) (Secretary)	

Electronic Measuring Equipment Subcommittee, ETDC 24 : 2

<i>Convener</i>	
DR RAM PARSHAD	National Physical Laboratory (CSIR), New Delhi
<i>Members</i>	
SHRI T. N. GHOSH (Alternate to Dr Ram Parshad)	
SHRI N. BALASUNDRAM	Eastern Electronics, Faridabad
LT-COL R. C. DHINGRA	Directorate General of Inspection (Ministry of Defence)
MAJ G. R. MAHADEVAN (Alternate)	
SHRI N. GANESAN	Associated Instrument Manufacturers (India) Pvt Ltd, New Delhi
SHRI D. PAHWA (Alternate)	
SHRI P. K. MANTRI	Toshniwal Bros Pvt Ltd, Ajmer
SHRI G. SUBBA RAO (Alternate)	

(Continued on page 9)

Indian Standard

SPECIFICATION FOR DIRECT READING POINTER INDICATOR TYPE AC/DC ELECTRONIC VOLTMETER

PART III UP TO 300 MHz

0. FOREWORD

0.1 This Indian Standard (Part III) was adopted by the Indian Standards Institution on 7 December 1972, after the draft finalized by the Electronic Equipment Sectional Committee had been approved by the Electrotechnical Division Council.

0.2 The provisions of this standard apply to the complete apparatus and not to component parts thereof.

0.3 This standard (Part III) should be used in conjunction with IS : 6767 (Part I)-1972*.

0.4 Assistance has been derived from the following while preparing this standard:

IEC Pub 217 Electronic voltmeter. International Electrotechnical Commission.

BS 4205 : 1967 Specification for electronic voltmeters. British Standards Institution.

0.5 This standard is one of a series of Indian Standards on electronic measuring equipment. A list of standards so far published in this series is given on Page 10.

0.6 For the purpose of deciding whether a particular requirement of this standard is complied with, the final value, observed or calculated, expressing the result of a test, shall be rounded off in accordance with IS : 2-1960†. The number of significant places retained in the rounded off value should be the same as that of the specified value in this standard.

*Specification for direct reading pointer indicator type ac/dc electronic voltmeter: Part I Methods of measurements.

†Rules for rounding off numerical values (*revised*).

1. SCOPE

1.1 This standard (Part III) lays down the minimum performance requirements and characteristics to be specified for direct reading pointer indicator type ac/dc electronic voltmeter operating up to 300 MHz.

1.2 The climatic and mechanical durability requirements for these voltmeters are under consideration.

2. TERMINOLOGY

2.1 For the purpose of this standard the definitions and explanation of terms specified in 2 of IS : 3437-1972* shall apply.

3. CONSTRUCTIONAL REQUIREMENTS

3.1 The provisions of 3 of IS : 3437-1972* shall apply as modified by 3.1.1.

3.1.1 The panel-meter shall have a minimum 100 divisions, each division being of 0.8 mm width.

4. WORKMANSHIP AND FINISH

4.1 The provisions of 5 of IS : 3437-1972* shall apply.

5. CONTROLS AND ADJUSTMENTS

5.1 The provisions of 6 of IS : 3437-1972* shall apply.

6. POWER SUPPLY

6.1 The provisions of 7 of IS : 3437-1972* shall apply.

7. CHARACTERISTICS OF ELECTRONIC VOLTMETERS

7.1 Basic Characteristics

7.1.1 Accuracy Class — The accuracy classes of 1.5 and 2.5 are covered in this standard (see also 2.8.2 and 4.1 of IS : 3437-1972*).

7.1.2 Permissible Intrinsic Error — The permissible intrinsic error under reference conditions and between the limits of the effective range shall be ± 1.5 percent or ± 2.5 percent.

7.1.3 Voltage Range

7.1.3.1 DC voltage range — The instrument shall be capable of measuring dc voltages from 1 V full scale deflection up to 300 V full scale deflection in suitable ranges.

*General requirements for direct reading pointer indicator type electronic voltmeter (first revision).

7.1.3.2 AC voltage range — The instrument shall be capable of measuring ac voltages from 1 V full scale deflection up to 300 V full scale deflection in suitable ranges.

7.1.3.3 The instrument may also be calibrated in decibels and a separate dB scale shall be provided.

7.1.3.4 The ratings of effective ranges shall preferably be chosen from the following basic series or their decimal multiples or factors (see 3.4.1 of IS : 3437-1972*):

- a) 1, $\sqrt{10}$, 10 (for voltage and dB); and
- b) 1, 3, 10 (for voltage).

7.1.4 Calibration — The calibration shall be for rms values. If it is otherwise, such as peak, average or logarithmic, it shall be indicated on the instrument.

7.1.5 Type of Input — The type of input shall be single ended. Other types of input such as symmetrical or differential may be provided.

7.1.6 Input Admittance|Impedance

7.1.6.1 For dc ranges — Input resistance shall be at least 1 000 megohms.

7.1.6.2 For ac ranges — The ac input impedance expressed in terms of its equivalent parallel, resistive and reactive components at 1 MHz shall be as follows:

Input resistance — 1 megohm, *Min*

Input capacitance — 3 pF, *Max*

7.1.7 Warm-Up Period — The time taken by the instrument to attain good stability within the intrinsic error specified shall not be greater than 15 minutes.

7.2 Other Characteristics

7.2.1 Deviation of Electrical Zero — The permissible shift of deviation of electrical zero for the two classes of voltmeters covered by the standard (see 7.1.2) shall be as specified in Table 1 for various influence quantities.

7.2.2 Variation in Indication — The permissible variation in indication for the two classes of voltmeter covered by the standard (see 7.1.1) shall be as specified in Table 2 for various conditions.

*General requirements for direct reading pointer indicator type electronic voltmeter (first revision).

TABLE 1 SHIFT OR DEVIATION OF ELECTRICAL ZERO

(Clause 7.2.1)

SL. No.	INFLUENCING QUANTITY	REF TO CL No. OF IS : 6767 (PART I)-1972*	SHIFT OR DEVIATION OF ELECTRICAL ZERO	
			Class Index 1.5	Class Index 2.5
(1)	(2)	(3)	(4)	(5)
			percent	percent
i)	Ambient temperature change	5.6	± 1.5	± 2.5
ii)	Supply voltage change:	5.5		
	a) First sudden 5 percent variation (maximum shift)		± 0.75	± 1.25
	b) Second sudden 5 percent variation (final shift)		± 1.5	± 2.5
iii)	Fluctuation	5.1	± 0.75	± 1.25
iv)	Drift	5.2	± 1.5	± 2.5
v)	Applied input voltage	5.3		
vi)	Change of range	5.4	± 1.5	± 2.5

NOTE — The deviation of electrical zero shall be expressed as percentage of the total length.

*Specification for direct reading pointer indicator type ac/dc electronic voltmeter: Part I Methods of measurements.

7.2.3 Damping — The damping of a voltmeter characterized by its overshoot and settling time shall be measured in accordance with 9 of IS : 6767 (Part I)-1972* and shall specify the following requirements:

- Overshoot** — Under the test conditions, the overshoot shall not exceed the upper limit of effective range.
- Settling time** — Under the test conditions the time required for the index to attain its final steady position within 1.5 percent of the upper limit of the effective range shall not exceed 4 seconds.

7.2.4 Errors Due to Overload

7.2.4.1 When subjected to the continuous overload test in accordance with 10.1 of IS : 6767 (Part I)-1972*, the voltmeter shall comply with the requirements of its specified accuracy.

*Specification for direct reading pointer indicator type ac/dc electronic voltmeters Part I Methods of measurements.

TABLE 2 VARIATION INDICATION

(Clause 7.2.2)

SL No.	INFLUENCING QUANTITY	REF TO CL No. OF IS : 6767 (PART I)-1972*	VARIATION IN INDICATION	
			Class Index 1.5	Class Index 2.5
(1)	(2)	(3)	(4) percent	(5) percent
i)	Position	6.1	± 1.5	± 2.5
ii)	Ambient temperature change	6.2	± 1.5	± 2.5
iii)	External magnetic field:	11		
	a) When specified by the manufacturer		± 1.5	± 2.5
	b) At 0.5 millitesla (when not specified by the manufacturer)		± 3.0	± 3.0
iv)	External electric field	12	Under consideration	
v)	External RF electromagnetic field	13	Under consideration	
vi)	Supply voltage change:	6.3		
	a) First sudden 5 percent variation (maximum variation)		± 0.75	± 1.25
	b) Second sudden 5 percent variation (final variation)		± 0.75	± 1.25
vii)	Super-imposed input voltage (see Note):	6.4		
	a) On ac, 1000 times dc voltage being applied		± 1.5	± 2.5
	b) On dc, 1.2 times ac voltage being applied		± 1.5	± 2.5
viii)	Frequency of measured voltage:	8		
	a) Over 40 Hz to 30 MHz		± 1.5	± 2.5
	b) Over 20 Hz to 300 MHz		± 3.0	± 5.0
ix)	Fluctuation (random and spurious variation)	6.5	± 0.75	± 1.25
x)	Drift	6.6	± 1.5	± 2.5

NOTE — In either of the cases the maximum voltage applied shall not exceed insulation voltage of the instrument.

*Specification for direct reading pointer indicator type ac/dc electronic voltmeter: Part I Methods of measurements.

IS : 6767 (Part III) - 1972

7.2.4.2 When subjected the test for overload for short duration in accordance with **10.2** of IS : 6767 (Part I)-1972*, the voltmeter shall comply with the requirements of its specified accuracy.

8. MARKING

8.1 The provisions of **8** of IS : 3437-1972† shall apply in addition to **8.1.1**.

8.1.1 The instrument may also be marked with the ISI Certification Mark.

NOTE—The use of the ISI Certification Mark is governed by the provisions of the Indian Standards Institution (Certification Marks) Act and the Rules and Regulations made thereunder. The ISI Mark on products covered by an Indian Standard conveys the assurance that they have been produced to comply with the requirements of that standard under a well-defined system of inspection, testing and quality control which is devised and supervised by ISI and operated by the producer. ISI marked products are also continuously checked by ISI for conformity to that standard as a further safeguard. Details of conditions under which a licence for the use of the ISI Certification Mark may be granted to manufacturers or processors, may be obtained from the Indian Standards Institution.

9. TECHNICAL DOCUMENTATION TO BE SUPPLIED WITH THE INSTRUMENT

9.1 The provisions of **9** of IS : 3437-1972† shall apply.

*Specification for direct reading pointer indicator type ac/dc electronic voltmeter: Part I Methods of measurements.

†General requirements for direct reading pointer indicator type electronic voltmeter (first revision).

(Continued from page 2)

Members

SHRI M. N. MATHUR
SHRI J. S. MONGA
SHRI S. B. NIRODI
SHRI D. V. PETKAR
SHRI D. V. S. RAJU

SHRI C. S. RANGAN

SHRI M. T. SRIVATSA (Alternate)

SHRI S. RANGARAJAN
RESEARCH ENGINEER
SHRI G. C. SAXENA

Representing

Telecommunication Research Centre (P & T Board)
Radiola Corporation, New Delhi
Philips (India) Ltd, Calcutta
Bhabha Atomic Research Centre, Bombay
Electronic & Industrial Instruments Co Pvt Ltd,
Hyderabad
National Aeronautical Laboratory (CSIR),
Bangalore

Bharat Electronics Ltd, Bangalore
Directorate General of All India Radio, New Delhi
Electronics Corporation of India Ltd, Hyderabad

INDIAN STANDARDS
ON
ELECTRONIC MEASURING EQUIPMENT

IS:

- 2320-1963 Methods of measurements for amplitude modulated radio frequency signal generators (30 kc/s to 30 Mc/s)
- 2321-1963 Requirements for general purpose amplitude modulated radio frequency signal generators (30 kc/s to 30 Mc/s)
- 2711-1966 Direct reading pH meters (*revised*)
- 3437-1972 General requirements for direct reading pointer indicator type electronic voltmeter (*first revision*)
- 3886-1966 Minimum requirements for general purpose audio frequency signal generators (30 c/s to 30 kc/s)
- 3915-1966 Methods of measurements on audio frequency signal generators (30 c/s to 30 kc/s)
- 4309-1967 Methods of measurement on direct reading pH meters
- 4330-1967 Methods of measurement on cathode-ray oscilloscope (dc to 10 Mc/s)
- 6700-1972 Requirements for general purpose cathode-ray oscilloscope
- 6756-1972 Technical documentation to be supplied with electronic measuring equipment
- 6767 (Part I)-1972 Direct reading pointer indicator type ac/dc electronic voltmeter: Part I Methods of measurements
- 6767 (Part II)-1972 Direct reading pointer indicator type ac/dc electronic voltmeter: Part II Up to 30 MHz
- 6767 (Part III)-1972 Direct reading pointer indicator type ac/dc electronic voltmeter: Part III Up to 300 MHz
- 6769 (Part I)-1972 Direct reading pointer indicator type ac electronic millivoltmeter: Part I Methods of measurements
- 6769 (Part II)-1972 Direct reading pointer indicator type ac electronic millivoltmeter: Part II With a frequency range of 2 Hz to 1 MHz
- 6804-1972 Glass electrodes for direct reading pH meters

PUBLICATIONS OF INDIAN STANDARDS INSTITUTION

INDIAN STANDARDS

Over 7 000 Indian Standards covering various subjects have been issued so far. Of these, the standards belonging to the Electrotechnical Group fall under the following categories:

Acoustics and acoustical measurement	Instrument transformers
Automobile electrical equipment	Insulating materials
Batteries	Insulators and accessories
Cinematographic equipment	Integrating meters
Conductors and cables	Lamps and lamp accessories
Domestic electrical appliances	Lifts and escalators
Electric welding equipment	Lightning arresters
Electrical installations, codes of practice	Motors and generators
Electrical instruments	Nomenclature and symbols
Electron tubes and valves	Power capacitors
Electronic components	Power converters
Electronic equipment	Relays
Environmental testing procedures	Rotating machinery
Fans	Semiconductor devices
Flameproof electrical equipment	Switchgear and controlgear
High voltage techniques	Transformers and reactors
Illuminating engineering	Winding wires
	Wiring accessories
	Unclassified

OTHER PUBLICATIONS

ISI Bulletin (Published Every Month)

Single Copy	Rs 3-00
Annual Subscription	Rs 25-00

Standards : Monthly Additions

Single Copy	Rs 0-30
Annual Subscription	Rs 3-00

Annual Reports (from 1948-49 Onwards) ... Rs 2-00 to 5-00

ISI Handbook, 1972 ... Rs 20-00

INDIAN STANDARDS INSTITUTION

Manak Bhavan, 9 Bahadur Shah Zafar Marg, New Delhi 1

Telephone : 27 01 31 (20 lines)

Telegrams : Manaksanstha

Branch Offices:

	Telephone
'Sadhna', Nurmohamed Shaikh Marg, Khanpur	Ahmedabad 1 2 03 91
F Block, Unity Bldg, Narasimharaja Square	Bangalore 2 2 76 49
534 Sardar Vallabhbhai Patel Road	Bombay 7 35 69 44
5 Chowringhee Approach	Calcutta 13 23-08 02
5-9-201/2-A (First Floor), Chirag Ali Lane	Hyderabad 1 3 44 35
117/418 B Sarvodaya Nagar	Kanpur 5 82 72
54 General Patters Road	Madras 2 8 72 78